

Technical Data Sheet

Article-No.: **081312**
 Description: **SOLID SATETY High Grip black**
Nitrile examination glove
black, non sterile, powder free



THICKNESS		
cuff	palm	fingertips
0,10 mm	0,17 mm	0,20 mm

PRODUCT DESCRIPTION							
material	<input type="checkbox"/> Latex	<input checked="" type="checkbox"/> Nitrile	<input type="checkbox"/> Vinyl	<input type="checkbox"/> Vinyl-Nitrile-mixture	<input type="checkbox"/> Polyethylene (PE)	<input type="checkbox"/> TPE	<input type="checkbox"/> cotton
colour	<input type="checkbox"/> white	<input type="checkbox"/> blue	<input checked="" type="checkbox"/> black	<input type="checkbox"/> mint	<input type="checkbox"/> purple	<input type="checkbox"/> mix	<input type="checkbox"/> apple-green
characteristics	<input type="checkbox"/> prepowdered	<input checked="" type="checkbox"/> powderfree	<input type="checkbox"/> sterile	<input checked="" type="checkbox"/> non sterile	<input checked="" type="checkbox"/> ambidextrous	<input type="checkbox"/> fits hand-specific	<input type="checkbox"/> biodegradable
surface	<input checked="" type="checkbox"/> full textured	<input type="checkbox"/> finger textured	<input type="checkbox"/> not textured	<input type="checkbox"/> embossed	<input checked="" type="checkbox"/> double chlorinated		

SIZES							
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)	XXXL (11-12)
width	-	80 ± 10 mm	95 ± 10 mm	110 ± 10 mm	115 ± 10 mm	125 ± 10 mm	-
length	-	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	-

REGULATORY AFFAIRS							
PPE-Regulation (EU) 2016/425	<input type="checkbox"/> Category I	<input type="checkbox"/> Category II	<input checked="" type="checkbox"/> Category III	<input type="checkbox"/> no PPE-article			CE
MD-Regulation (EU) 2017/745	<input checked="" type="checkbox"/> Class I	<input type="checkbox"/> Class II a	<input type="checkbox"/> Class III	<input type="checkbox"/> sterile	<input type="checkbox"/> measuring function	<input type="checkbox"/> no medical device	
Food Contact (EG) 1935/2004	<input checked="" type="checkbox"/> acidic foods	<input checked="" type="checkbox"/> aqueous foods	<input checked="" type="checkbox"/> fatty foods	<input checked="" type="checkbox"/> alcoholic foods	<input checked="" type="checkbox"/> dry foods	<input type="checkbox"/> not approved for food-contact	RF

STANDARDISATION						
EN 388 Mechanical Risks	abrasion resistance	blade cut resistance Coupe-Test	tear resistance	puncture resistance	blade cut resistance TDM-Test	impact test
Level	not applicable					

EN 374-1 Chemical Risks	chemical	code letter	level	permeation time	degradation	ISO 374-1/Type B
	Sodium hydroxide 40%	K	6	> 480 min	-5,8 %	
	Hydrogen Peroxide 30%	P	6	> 480 min	17,2 %	
EN 374-4 Degradation	Formaldehyde 37%	T	6	> 480 min	26,1 %	KPT
EN 374-5 microorganism tightness	The glove is tight against microorganisms (viral, bacteria and fungi). Test according to ISO 16604 - method B.					EN ISO 374-5:2016 VIRIUS

EN 420 protective gloves	The glove meets the requirements according to EN 420					
--------------------------	--	--	--	--	--	--

EN 455 medical gloves for single use	The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4					EN 455
--------------------------------------	--	--	--	--	--	--------

EN 455-1 freedom from holes	The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I)					AQL 1.5
-----------------------------	--	--	--	--	--	---------

ISO 18889 Pesticide application	not applicable					
---------------------------------	----------------	--	--	--	--	--

Technical Data Sheet

Article-No.: **081312**

Description: **SOLID SATETY High Grip black**
Nitrile examination glove
black, non sterile, powder free



LOGISTIC DATA SUBPACKING		
generell information		
material	carton	
pieces per subpacking	100	
GTIN subpacking size XS	-	
GTIN subpacking size S	4044941719894	
GTIN subpacking size M	4044941719917	
GTIN subpacking size L	4044941719931	
GTIN subpacking size XL	4044941719955	
GTIN subpacking size XXL	4044941719979	
GTIN subpacking size XXXL	-	
PZN subpacking size XS	-	
PZN subpacking size S	-	
PZN subpacking size M	-	
PZN subpacking size L	-	
PZN subpacking size XL	-	
PZN subpacking size XXL	-	
PZN subpacking size XXXL	-	
measures & size		
length	239 mm	
width	125 mm	
height	85 mm	
weights		
size	net weight	gross weight
XS	-	-
S	590 g	650 g
M	650 g	710 g
L	740 g	800 g
XL	770 g	830 g
XXL	800 g	860 g
XXXL	-	-

LOGISTIC DATA PALETTE		
general information		
kind of palett	euro-palette	
measures & size		
cartons per layer	6	
layers per palette	7	
height of the palette	190 cm	
weights		
size	net weight	gross weight
XS	-	-
S	294 kg	319 kg
M	319 kg	344 kg
L	357 kg	382 kg
XL	370 kg	395 kg
XXL	382 kg	407 kg
XXXL	-	-

LOGISTIC DATA OUTER PACKING		
generell information		
material	carton	
subpackings per outer packing	10	
GTIN outer packing size XS	-	
GTIN outer packing size S	4044941719900	
GTIN outer packing size M	4044941719924	
GTIN outer packing size L	4044941719948	
GTIN outer packing size XL	4044941719962	
GTIN outer packing size XXL	4044941719986	
GTIN outer packing size XXXL	-	
PZN outer packing size XS	-	
PZN outer packing size S	-	
PZN outer packing size M	-	
PZN outer packing size L	-	
PZN outer packing size XL	-	
PZN outer packing size XXL	-	
PZN outer packing size XXXL	-	
measures & size		
length	440 mm	
width	260 mm	
height	250 mm	
weights		
size	net weight	gross weight
XS	-	-
S	6.500 g	7.000 g
M	7.100 g	7.600 g
L	8.000 g	8.500 g
XL	8.300 g	8.800 g
XXL	8.600 g	9.100 g
XXXL	-	-

Technical Data Sheet

Article-No.: **081312**
 Description: **SOLID SATETY High Grip black**
Nitrile examination glove
black, non sterile, powder free



WARNINGS AND SAFETY INFORMATION

storage / expiry date	Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.
use and control	Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical. Before use, the gloves must be checked for holes or damage.
disposal	Used gloves must be disposed of after contact with chemicals in accordance with the disposal regulations for the chemical and the regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.
disinfection	Disinfection is not intended for these gloves and is the responsibility of the user.
warnings/ allergy information	Protective gloves are intended for single use only. This product contains dithiocarbamates, which may cause allergic reactions
donning and doffing instructions	

Technical Data Sheet

Article-No.: **081312**

Description: **SOLID SATETY High Grip black**
Nitrile examination glove
black, non sterile, powder free

date 03.09.2024



changes and errors excepted